

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 27 2011

Datamaster cdm S/N 130131

Supervisor/Operator Performing the Verification Procedure:

Name Shayne Calt ID# 4559 Date 6-29-11

A Agency AST Phone # 733-2258

Instrument Location Talkeetna AST

B Alco S/N D119594 Target Value 0.078 High Pressure 200
8MB 8/11/11

Alco Test Value Average .075 .075
 1st Alco 2nd Alco

Signature [Signature]

(OVER)

(Do Not write in the area below)

8MB
8/11/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz 8/18/11
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this 18 day of Aug, 2010 2011

Nikki Roth (Notary Seal Stamp)
 Daniel L. Nicholas, Nikki Roth
 Notary Public, State of Alaska
 Commission Expires with Office



